

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE**CERTIFICAT D'ESSAI OC**

Product
Produit

Switch Mode Power Supply

Name and address of the applicant
Nom et adresse du demandeur

RAM TECHNOLOGIES L L C
UNIT 12
29 SOUNDVIEW RD
GUILFORD CT 06437 UNITED STATES

Name and address of the manufacturer
Nom et adresse du fabricant

RAM TECHNOLOGIES L L C
UNIT 12
29 SOUNDVIEW RD
GUILFORD CT 06437 UNITED STATES

Name and address of the factory
Nom et adresse de l'usine

RAM TECHNOLOGIES L L C
UNIT 12
29 SOUNDVIEW RD
GUILFORD CT 06437 UNITED STATES

Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{eme} page

Additional Information on page 2
See Page 2

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

None

Trademark (if any)
Marque de fabrique (si elle existe)
Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais
constructeur

Model / Type Ref.
Ref. De type

PFC800PCX, PFC600PCX

Additional information (if necessary may also be
reported on page 2)
Les informations complémentaires (si nécessaire,,
peuvent être indiqués sur la 2^{eme} page

National Differences specified in the CB Test Report.
 Additional Information on page 2

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

IEC 60601-1(ed.3)

As shown in the Test Report Ref. No. which forms part
of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

E191947-D1000-1/A0/C0-ULCB issued on 2016-05-13

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/nbcnames

Date: 2016-05-19

Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

US-27550-UL

Ratings:

Model PFC600PCX

Input Ratings:

100-240 Vac, 50-60 Hz, 7.5 A max

Output Ratings:

550W Continuous / 600W Peak

Max output current: +3.3 Vdc, 24 A; +5 Vdc, 24 A; +12 Vdc, 45.8A/50A Peak; -12 Vdc, 0.5 A; 5 Vdc(5VSB), 3A/3.5AP;

Total output of +12 V, +5 V, +3.3 V, -12 V & 5 VSB, shall not exceed 550 Watts Continuous / 600W peak

Max Output from 3.3V and 5V Output shall not exceed 140Watts.

Model PFC800PCX

Input Ratings:

100-240 Vac, 50-60 Hz, 10 A max

Output Ratings:

750W Continuous / 800W Peak

Max output current: +3.3 Vdc, 24 A; +5 Vdc, 24 A; +12 Vdc, 62.5A/67A Peak; -12 Vdc, 0.5 A; 5 Vdc(5VSB), 3A/3.5AP;

Total output of +12 V, +5 V, +3.3 V, -12 V & 5 VSB, shall not exceed 750 Watts Continuous / 800W peak

Max Output from 3.3V and 5V Output shall not exceed 140Watts.

Additional Information:

Risk Management was not included in this investigation.

Additional information (if necessary)

Information complémentaire (si nécessaire)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2016-05-19

Signature:

Jolanta M. Wroblewska