

IEC SYSTEM FOR CONFORMITY TESTING AND  
CERTIFICATION OF ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME

SYSTEME CEI D'ESSAIS DE CONFORMITE ET DE CERTIFICATION  
DES EQUIPEMENTS ELECTRIQUES (IECEE)  
METHODE OC

## CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product  
*Produit*

Switch Mode Power Supply

Name and address of the applicant  
*Nom et adresse du demandeur*

RAM TECHNOLOGIES L L C  
UNIT 12 29 SOUNDVIEW RD  
GUILFORD CT 06437, USA

Name and address of the manufacturer  
*Nom et adresse du fabricant*

RAM TECHNOLOGIES L L C  
UNIT 12 29 SOUNDVIEW RD  
GUILFORD CT 06437, USA

Name and address of the factory  
*Nom et adresse de l'usine*

RAM TECHNOLOGIES L L C  
UNIT 12 29 SOUNDVIEW RD  
GUILFORD CT 06437, USA

Rating and principal characteristics  
*Valeurs nominales et caractéristiques principales*

For ratings see the second page of this certificate.

Trademark (if any)  
*Marque de fabrique (si elle existe)*

**RAM Technologies**

Model / Type Ref.  
*Ref. de type*

PFC600PCX

Additional information (if necessary)  
*Information complémentaire (si nécessaire)*

This CB Test Report comprises 7 enclosures.

A sample of the product was tested and found  
to be in conformity with  
*Un échantillon de ce produit a été essayé et a été  
considéré conforme à la*

as shown in the Test Report Ref. No.  
which forms part of this Certificate  
*comme indiqué dans le Rapport d'essais numéro  
de référence qui constitue partie de ce Certificat*

### PUBLICATION

**IEC 60601-1 (1988) Second Edition,  
with Amendment No. 1 (1991) and No. 2 (1995)** with the exception of:  
Clause 36, Electromagnetic Compatibility, Clause 48, Biocompatibility and  
Clause 52.1, Programmable Electronic Systems. Inclusive of CENELEC Common  
Modifications. See Test Report for National Differences.

E191947-A3-CB-1

This CB Test Certificate is issued by the National Certification Body  
*Ce Certificat d'essai OC est établi par l'Organisme National de Certification*



**Underwriters  
Laboratories**

Underwriters Laboratories Inc. / GMA Certification Department, US  
333 Pfingsten Road, Northbrook, IL 60062-2096  
United States of America  
TEL INT\* +1 847 664 3008, FAX INT\* +1 847 313 3008  
email: jolanta.m.wroblewska@us.ul.com

Date: Issued: 2010 January 26

Signature:

Jolanta M. Wroblewska

IEC SYSTEM FOR CONFORMITY TESTING AND  
CERTIFICATION OF ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME

SYSTEME CEI D'ESSAIS DE CONFORMITE ET DE CERTIFICATION  
DES EQUIPEMENTS ELECTRIQUES (IECEE)  
METHODE OC

## CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Input Ratings:  
100-240 Vac, 50-60 Hz, 7.5 A max

Output Ratings:  
550W Continuous / 600W Peak

Max output current: +3.3 Vdc, 24 A; +5 Vdc, 24 A; +12 Vdc, 45.8 A/50A Peak; -12 Vdc, 0.5 A; 5 Vdc(5VSB), 3A/3.5AP;

Total output of +12 V, +5 V, +3.3 V, -12 V & 5 VSB, shall not exceed 550 Watts Continuous / 600W peak

Max Output from 3.3V and 5V Output shall not exceed 140 Watts.

Additional information (if necessary)  
*Information complémentaire (si nécessaire)*

This CB Test Certificate is issued by the National Certification Body  
*Ce Certificat d'essai OC est établi par l'Organisme National de Certification*




**Underwriters  
Laboratories**

Date: Issued: 2010 January 26

Underwriters Laboratories Inc. / GMA Certification Department, US  
333 Pfingsten Road, Northbrook, IL 60062-2096  
United States of America  
TEL INT\* +1 847 664 3008 FAX INT\* +1 847 313 3008  
email: [jolanta.m.wroblewska@us.ul.com](mailto:jolanta.m.wroblewska@us.ul.com)

Signature:



Jolanta M. Wroblewska