

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC

**CB TEST CERTIFICATE****CERTIFICAT D'ESSAI OC**

Product  
Produit

Switch Mode Power Supply

Name and address of the applicant  
Nom et adresse du demandeur

RAM TECHNOLOGIES L L C  
UNIT 12, 29 SOUNDVIEW RD  
GUILFORD CT 06437, USA

Name and address of the manufacturer  
Nom et adresse du fabricant

RAM TECHNOLOGIES L L C  
UNIT 12, 29 SOUNDVIEW RD  
GUILFORD CT 06437, USA

Name and address of the factory  
Nom et adresse de l'usine

RAM TECHNOLOGIES L L C  
UNIT 12, 29 SOUNDVIEW RD GUILFORD CT 06437  
USA

Note: When more than one factory, please report on page 2  
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2<sup>ème</sup> page

Additional Information on page 2

Ratings and principal characteristics  
Valeurs nominales et caractéristiques principales

See Page 2

Trademark (if any)  
Marque de fabrique (si elle existe)



Type of Manufacturer's Testing Laboratories used  
Type de programme du laboratoire d'essais  
constructeur

Model / Type Ref.  
Ref. De type

PFC230 PC, PFC230 PCX, PFC280 PCX, PFC310 PCX,  
PFC175 SFX, PFC200 SFX, and PFC400PCX

Additional information (if necessary may also be  
reported on page 2)  
Les informations complémentaires (si nécessaire,,  
peuvent être indiqués sur la 2<sup>ème</sup> page

National Differences specified in the CB Test Report. The Risk  
Management requirements of the standard were not addressed.

Additional Information on page 2

A sample of the product was tested and found  
to be in conformity with  
Un échantillon de ce produit a été essayé et a été  
considéré conforme à la

IEC 60601-1(ed.3)

As shown in the Test Report Ref. No. which forms  
part of this Certificate  
Comme indiqué dans le Rapport d'essais numéro de  
référence qui constitue partie de ce Certificat

11ME06231 issued on 2011-11-30

This CB Test Certificate is issued by the National Certification Body

Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2016-10-03

Signature:

Original Issue Date: 2011-12-01

Jolanta M. Wroblewska



Ratings:

Input Ratings:

Models PFC230 PC, PFC230 PCX:  
100-240 VAC, 47-63 Hz, 3.75 A max

Model PFC280 PCX:  
100-240 VAC, 47-63 Hz, 4.0 A max

Model PFC310 PCX:  
100-240 VAC, 47-63 Hz, 4.75 A max

Model PFC175 SFX:  
100-240 VAC, 50-60 Hz, 2.5 A max

Model PFC200 SFX:  
100-240 VAC, 50-60 Hz, 3 A max

Model PFC400PCX:  
100-240 Vac, 50-60 Hz, 5.75 A max

For Model Output Ratings see Test Report.

Additional Information:

The CB Test Certificate was corrected to address exclusion of Risk Management requirements.

**Additional information (if necessary)**

**Information complémentaire (si nécessaire)**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2016-10-03  
Original Issue Date: 2011-12-01

Signature:

Jolanta M. Wroblewska